

Accreditation



The Deutsche Akkreditierungsstelle attests with this **Accreditation Certificate** that the calibration laboratory

Göttfert Werkstoff - Prüfmaschinen GmbH
Siemensstraße 2, 74722 Buchen Odenwald

meets the requirements according to DIN EN ISO/IEC 17025:2018 for the conformity assessment activities listed in the annex to this certificate. This includes additional existing legal and normative requirements for the calibration laboratory, including those in relevant sectoral schemes, provided they are explicitly confirmed in the annex to this certificate.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the principles of DIN EN ISO 9001.

This accreditation was issued in accordance with Art. 5 Para. 1 Sentence 2 of Regulation (EC) 765/2008, after an accreditation procedure was carried out in compliance with the minimum requirements of DIN EN ISO/IEC 17011 and on the basis of a review and decision of the appointed accreditation committees.

This accreditation certificate only applies in connection with the notices of 21.05.2025 with accreditation number D-K-17291-02.


It consists of this cover sheet, the reverse side of the cover sheet and the following annex with a total of 3 pages.

Registration number of the accreditation certificate: **D-K-17291-02-00**

Berlin, 21.05.2025

Dipl.-Wirtsch.-Ing. (BA) Tim Harnisch
Head of Technical Unit

Translation issued:
21.05.2025



Dipl.-Wirtsch.-Ing. (BA) Tim Harnisch
Head of Technical Unit

The certificate together with the annex reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH (www.dakks.de).

This document is a translation. The definitive version is the original German accreditation certificate.

See notes overleaf

Deutsche Akkreditierungsstelle GmbH

Office Berlin
Spittelmarkt 10
10117 Berlin

Office Frankfurt am Main
Europa-Allee 52
60327 Frankfurt am Main

Office Braunschweig
Bundesallee 100
38116 Braunschweig

The Deutsche Akkreditierungsstelle GmbH (DAkKS) is the entrusted national accreditation body of the Federal Republic of Germany according to § 8 section 1 AkkStelleG in conjunction with § 1 section 1 AkkStelleGBV. DAkKS is designated as the national accreditation authority by Germany according to Art. 4 Para. 4 of Regulation (EC) 765/2008 and clause 4.7 of DIN EN ISO/IEC 17000.

Pursuant to Art. 11 section 2 of Regulation (EC) 765/2008, the accreditation certificate shall be recognised as equivalent by the national authorities within the scope of this Regulation as well as by the WTO member states that have committed themselves in bilateral or multilateral mutual agreements to recognise the certificates of accreditation bodies that are members of ILAC or IAF as equivalent.

DAkKS is a signatory to the multilateral agreements for mutual recognition of the European co-operation for Accreditation (EA), International Accreditation Forum (IAF) and International Laboratory Accreditation Co-operation (ILAC).

The up-to-date state of membership can be retrieved from the following websites:

EA: www.european-accreditation.org

ILAC: www.ilac.org

IAF: www.iaf.nu

Deutsche Akkreditierungsstelle

Annex to the Accreditation Certificate D-K-17291-02-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 21.05.2025

Date of issue: 21.05.2025

Holder of accreditation certificate:

Göttfert Werkstoff - Prüfmaschinen GmbH
Siemensstraße 2, 74722 Buchen Odenwald

with the location

Göttfert Werkstoff - Prüfmaschinen GmbH
Siemensstraße 2, 74722 Buchen Odenwald

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and they conform to the principles of DIN EN ISO 9001.

Calibrations in the fields:

Mechanical Quantities

Material Testing Machines (MTM)

- **Velocity (MTM) ^{a)}**
- **Force (MTM) ^{a)}**
- **Extension (MTM) ^{a)}**
- **Temperature (MTM) ^{a)}**

^{a)} also on-site calibration

This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at <https://www.dakks.de>.

Abbreviations used: see last page

Page 1 of 3

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Accreditation Certificate D-K-17291-02-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Temperature (MTM) Temperature of temperature chambers (incl. temperature distribution)	50 °C to 299.99 °C		0.06 °C	Calibration of the PT100 sensors on the heating tapes of the channel
	300 °C to 500 °C		0.07 °C	
Force (MTM) Weight force	3 N to 220 N	GFT NORM 204:2025-03	0.06 %	Procedure according to comparison method with reference
Velocity (MTM)	0.1 cm/min to 60 cm/min	Combined measurand GFT NORM 205:2025-03 GFT NORM 206:2025-03	0.25 %	Combined measureand from procedure according to comparison method with reference
Extension (MTM) Position	1 mm to 100 mm	GFT NORM 206:2025-03	0.01 mm	Procedure according to comparison method with reference
Test stamp, diameter	1 mm to 50 mm	GFT NORM 200:2025-03	0.004 mm	Using calibrated outside micrometer / caliper gauge, measure corresponding size at various points
Test stamp, length	1 mm to 150 mm		0.05 mm	
Test channel, inner diameter	8 mm to 35 mm	GFT NORM 201:2025-03	0.003 mm	Measure with calibrated fine pointer at several points
Capillary, inner diameter	0.5 mm to 8 mm	GFT NORM 202:2025-03	0.003 mm	Measure with calibrated fine pointer/iron micrometer at several points
Capillary, length	1 mm to 50 mm		0.013 mm	

Valid from: 21.05.2025

Date of issue: 21.05.2025

Page 2 of 3

This document is a translation. The definitive version is the original German annex to the accreditation certificate.

Annex to the Accreditation Certificate D-K-17291-02-00

On-site Calibration

Calibration and Measurement Capabilities (CMC)					
Measurement quantity / Calibration item	Range		Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Temperature (MTM) Temperature of temperature chambers (incl. temperature distribution)	50 °C to	299.99 °C	GFT NORM 203:2025-03	0.06 °C	Calibration of the PT100 sensors on the heating tapes of the channel
	300 °C to	500 °C		0.07 °C	
Force (MTM) Weight force	3 N to	220 N	GFT NORM 204:2025-03	0.06 %	Procedure according to comparison method with reference
Velocity (MTM)	0.1 cm/min to	60 cm/min	Combined measurand GFT NORM 205:2025-03 GFT NORM 206:2025-03	0.25 %	Combined measurand from procedure according to comparison method with reference
Extension (MTM) Position	1 mm to	100 mm	GFT NORM 206:2025-03	0.01 mm	Procedure according to comparison method with reference
Test stamp, diameter	1 mm to	50 mm	GFT NORM 200:2025-03	0.004 mm	Using calibrated outside micrometer / caliper gauge, measure corresponding size at various points
Test stamp, length	1 mm to	150 mm		0.05 mm	
Test channel, inner diameter	8 mm to	35 mm	GFT NORM 201:2025-03	0.003 mm	Measure with calibrated fine pointer at several point

Abbreviations used:

CMC	Calibration and measurement capabilities
DIN	Deutsches Institut für Normung e.V. – German institute for standardization
EN	Europäische Norm – European Standard
GFT	internal calibration procedure of Göttfert Werkstoff - Prüfmaschinen GmbH
IEC	International Electrotechnical Commission
ISO	International Organization for Standardisation

Valid from: 21.05.2025

Date of issue: 21.05.2025

Page 3 of 3

This document is a translation. The definitive version is the original German annex to the accreditation certificate.